




Read the corresponding technical documentation for handling and safety reasons.

Running-in information



After installing a new deuterium lamp into your detector, ensure to allow the lamp a running-in time of approximately 24 hours.

For further details on installing the lamp, consult the technical documentation of the detector.