

Device support in KNAUER CDS / KNAUER device support in 3rd party software OpenLab® CDS EZChrom Edition / ClarityChrom® / Chromeleon™ / ChromGate® / PurityChrom® other software

The tables list the software package and driver version (in case multiple versions exist); **the version number listed is the first version that supports the respective device**. Newer software versions also include the driver. This does not apply to Chromeleon 6.8/7.x and PurityChrom 5.x/6.x.

Limitations are marked with superscript, e. g. ¹⁾. Limitations may affect device functionality, control interfaces or software versions. "n/a" means, that a device is not supported, "-" that no information is available. If a device is supported via different communication ports, for most devices the desired port has to be selected in the device directly. From version 7.4.1, ClarityChrom® *Prep* is discontinued; all its prep-functionality is now available in ClarityChrom® with KNAUER FRC license. For some time, ChromGate® was also available as "PLATINblue Edition" for the PLATINblue devices, while the standard version is named as "regular". From driver version 2825 all drivers are available in the regular version; the PLATINblue Edition was discontinued. ChromGate® was discontinued in 2013 at all. OpenLab® CDS EZChrom Edition version number lists Shared Service version and EZChrom version. Support for Chromeleon™ 6.8 and 7.x is listed separately in the Chromeleon™ column; if not mentioned, no driver exists for this version. The same applies for PurityChrom versions 5.x and 6.x; both listed in the PurityChrom column separately. KNAUER **OEM** devices are **not** supported by the regular KNAUER drivers. At the end of the document other software packages, e. g. KNAUER Mobile Control, are listed. The "Table of contents" includes links for jumping to the desired part of the document.

Drivers for KNAUER devices for Chromeleon™ 7.3.x, OpenLab® CDS ChemStation Edition or OpenLab® CDS (2.x) are not available.

Table of contents

Pumps	2	Mobile Control - KNAUER	25
Detectors	4	EuroChrom for Windows 3.05	26
Assistants	12	Other Software	27
Autosamplers	13	OpenLAB CDS (Agilent).....	27
Analog/Digital Converters	16	OpenLAB CDS ChemStation Edition (Agilent).....	27
Thermostats/Column Ovens	17	LabVIEW (National Instruments)	27
Electrical Valve Drives	18	HyStar™	27
Other Supported Instruments	20	Xcalibur™	27
Fraction Collectors	21		

Pumps	OpenLab® CDS EZChrom Edition CDS version and Driver version	ClarityChrom® CDS version and Driver version	Chromleon™ CDS version Driver version	ChromGate® CDS version and Driver version	PurityChrom® CDS and Driver version
AZURA® Pump P 8.1L	OLEE A.02.00 / A.04.06 Driver 2197	CC 8.7.0.107 Driver 6076	n/a	n/a	n/a
AZURA® Pump P 6.1L	OLEE A.01.05 / A.04.05 Driver 1022 ¹⁴⁾	CC 5.0.2 Driver 4095 ^{1, 15)}	Cm 7.2 Regular Class 3 driver 1024 ¹⁾	n/a	Version 6 R1-2 ¹⁾ Version 5.07.039 ¹⁾
AZURA® Pump P 2.1L	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.7 Driver 3078 ¹⁾	Cm 7.2 Regular Class 3 driver 1024 ¹⁾ Cm 6.8 Generic driver ²⁾	CG 3.3.2 Driver 2825 ³⁾	Version 6 R1-2 ¹⁾ Version 5.07.039 ^{1), 16)}
AZURA® Pump P 2.1S/P 4.1S ⁶⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.7 Driver 3078	Cm 7.2 Regular Class 3 driver 1056 ¹⁾ Cm 6.8 Generic driver ²⁾	CG 3.3.2 Driver 1787	Version 6 R1-2 Version 5.07.039
PLATINblue Pump P-1	OLEE A.01.05 / A.04.05 Driver 1017	CC 5.0.3 Driver 4102	Cm 6.8 SR10 Regular Class 3 driver	CG 3.3.2 PLATINblue Edition 1751; Regular 2825	n/a
Smartline Pump 1050 ⁴⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.6.5 Driver 1100	Cm 6.8 SR10 Regular Class 3 driver	CG 3.3.2 Driver 1751	Version 5.07.039
Smartline Pump 1000 ¹³⁾	n/a	CC 2.4.4	Cm 6.8 Generic driver ²⁾	CG 3.1	Version 5.07.039
Smartline Pump 100 ^{5), 6)}	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.6	Cm 6.8 Generic driver ²⁾	CG 3.3.1 Driver 1640	Version 5.07.039
Preparative Pump 1800 (also WellChrom K-1800) ¹³⁾	n/a	CC 2.6	n/a	CG 3.1	Version 5.07.039

Pumps (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver version	ClarityChrom® CDS version and Driver version	Chromleon™ CDS version Driver version	ChromGate® CDS version and Driver version	PurityChrom® CDS and Driver version
WellChrom Pump K-1001 ¹³⁾	n/a	CC 2.6	n/a	CG 2.55 ⁷⁾	Version 5.07.039
WellChrom Pump K-1000 ¹³⁾	n/a	CC 2.6	n/a	CG 2.55 (not supported in version 3.1) ⁸⁾	n/a
WellChrom Pump K-501 ⁹⁾	n/a	CC 2.6	n/a	CG 2.8	n/a
WellChrom Pump K-120	n/a	CC 2.6	n/a	CG 2.8	n/a
BlueShadow Pumps 10P/20P ⁶⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.7 Driver 3078	Cm 6.8 Generic driver ²⁾	CG 3.3.2 Driver 1787	Version 5.07.039
BlueShadow Pump 40P	OLEE A.01.05 / A.04.05 Driver 1022	CC 5.0.2 Driver 4095 ¹⁰⁾	n/a	n/a	Version 5.07.039
BlueShadow Pump 80P	OLEE A.01.05 / A.04.05 Driver 1022	CC 5.0.2 Driver 4095 ¹⁰⁾	Cm 6.8 Generic driver ²⁾	n/a	Version 6 R1-2 Version 5.07.039
Kontron Pumps 320/322/325	n/a	CC 3.0.5 ¹¹⁾	-	CG 3.1.7 Driver 446 ¹²⁾	n/a
Kontron Pumps 420/422/425	n/a	CC 3.0.5 ¹¹⁾	-	CG 3.1.7 Driver 446 ¹²⁾	Version 5.07.039
Kontron Pumps 520/522/525	n/a	CC 3.0.5 ¹¹⁾	-	CG 3.1.7 Driver 446 ¹²⁾	n/a

- 1) - Constant Pressure Mode / Isobar Mode is not supported
- 2) - generic drivers do not support all device features (e. g. no Isobar Mode); only control via RS-232; should also work with Chromleon™ 7.x
- 3) - support of RS-232 port from driver version 2844
- 4) - for control via RS-232 only baud rate of 115200 is supported
- 5) - can be controlled as pump K-120 only via RS-232 without pressure sensor in former software versions
- 6) - to set a static IP address a separate tool is required
- 7) - from ChromGate® version 3.1 only with firmware version 5.xx or higher; pumps with firmware up to version 4.XX has to be configured as K-1000
- 8) - from ChromGate® version 3.1 no pressure information; also K-1001 with firmware up to version 4.XX, no pressure trace
- 9) - from firmware version 1.23, model K-500 is not supported

- 10) - supported from Clarity version 4.X
- 11) - driver was developed by DataApex
- 12) - special RS-232 cable is required
- 13) - for full functionality the KNAUER RS-232 Y-cable is required (KNAUER part number A0755)
- 14) - support of switching SSVs during run from driver version 4.2.0.2161
- 15) - support of switching SSVs during run from driver version 6021
- 16) - only supported as isocratic pump in PurityChrom 5

Detectors [for all PDA or scanning UV detectors the 3D UV license option is required to acquire and review the scan (3D) data]	OpenLab® CDS EZChrom Edition CDS version and Driver version	ClarityChrom® CDS version and Driver version	Chromleon™ CDS version and Driver version	ChromGate® CDS version and Driver version	PurityChrom® CDS and Driver version
AZURA® Detector UVD 2.1L ¹⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.7 Driver 3078	Cm 7.2 Regular Class 3 driver 1024 Cm 6.8 Generic driver ⁵⁾	CG 3.3.2 Driver 2825	Version 6 R1-2 Version 5.08.034 ²⁷⁾
AZURA® Detector MWD 2.1L	OLEE A.01.05 / A.04.05 driver 2068	CC 5.0.5 Driver 5011	Cm 7.2 Regular Class 3 driver 1024	n/a	Version 6 R1-2 ¹⁵⁾ Version 5.07.067 ^{15), 27)}
AZURA® Detector UVD 2.1S ²⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.7 Driver 3055	Cm 7.2 Regular Class 3 driver 1056	CG 3.3.2 Driver 1787	Version 6 R1-2 Version 5.07.039 ²⁷⁾
AZURA® Detector DAD 6.1L ²³⁾	OLEE A.01.05 / A.04.05 Driver 2047	CC 5.0.5 Driver 4110	Cm 7.2 Regular Class 3 driver 1024	n/a	Version 6 R2-2 Version 5.07.102 ¹⁶⁾
AZURA® Detector DAD 2.1L ²³⁾	OLEE A.01.05 / A.04.05 Driver 2068	CC 5.0.5 Driver 5011	Cm 7.2 Regular Class 3 driver 1024	n/a	Version 6 R2-2 Version 5.07.102 ¹⁶⁾

Detectors (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver version	ClarityChrom® CDS version and Driver version	Chromeleon™ CDS version Driver version	ChromGate® CDS version and Driver version	PurityChrom® CDS and Driver version
AZURA® Detector RID 2.1L / RID 2.1L Highflow	OLEE A.01.05 / A.04.05 Driver 2092	CC 6.0.1 Driver 5037 (compatible with CC/CCP 5.x versions)	Cm 7.2 Regular Class 3 driver 1038	n/a	Version 6 R2-2 Version 5.08.024 ^{17), 18) 25)}
AZURA® Conductivity Monitor CM 2.1S ²⁾	OLEE A.01.05 / A.04.05 Driver 1022	CC 5.0.2 Driver 4095	n/a	Generic Driver June 2013 ⁵⁾	Version 6 R1-2 Version 5.07.039 ¹⁸⁾
Runge Mikron 81 Conductivity Monitor	n/a	CC 8.5, supported by manufacturer	n/a	n/a	Version 6 R1-2 Version 5.09.149 ¹⁸⁾
AZURA® ECD 2.1	n/a	Supported as Antec Decade Elite; CC 5.X, USB, LAN	Cm 7.2 Regular Class 3 driver	n/a	n/a
PLATINblue Detector MW-1 ³⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC6.1.0; Driver 5015	Cm 6.8 SR10 Regular Class 3 driver	CG 3.3.2 PLATINblue Edition 1751; Regular 2825	n/a
PLATINblue DAD PDA-1 ²³⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 5.0.2 Driver 4095 ⁴⁾	Cm 6.8 SR10 Regular Class 3 driver	CG 3.3.2; PLATINblue Edition 1751; Regular 2825	n/a
Smartline UV Detector 2500 ⁶⁾	n/a	CC 2.4.4	Cm 6.8 Generic driver ⁵⁾	CG 3.1	Version 5.07.039
Smartline UV Detector 2520 ^{3), 7)}	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.5 Driver 3033	n/a	CG 3.3.2 Driver 1751	n/a
Smartline UV Detector 2550 ^{3), 7), 23)}	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.6.5 Driver 1079	Cm 6.8 SR10 Regular Class 3 driver	CG 3.3.1 Driver 1640	n/a

Detectors (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromleon™ CDS version Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
Smartline UV Detector 2600 ²³⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.4.4	Cm 6.8 Generic driver ⁵⁾	CG 3.1	Version 5.07.039
Smartline UV Detector 200	n/a	CC 2.6	n/a	CG 3.1.7 Driver 1466	n/a
Smartline PDA Detector 2800 ^{9), 23)}	n/a	CC 2.4.4	n/a	CG 3.1.7 Driver 446	Version 5.07.039
Smartline PDA Detector 2850 ²³⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.6.5 Driver 1100	n/a	CG 3.3.2 Driver 1751	n/a
Smartline RI Detector 2300	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.4.4	n/a	CG 3.1	n/a
Smartline RI Detector 2400	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.4.4	n/a	CG 3.1	n/a
Smartline Conductivity Monitor S2900	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.7 Driver 3055	n/a	Generic Driver September 2010	n/a
WellChrom UV Detector K-2000 ⁸⁾	n/a	CC 2.6	n/a	CG 2.8	n/a
WellChrom UV Detector K-2001	n/a	CC 2.6	n/a	CG 2.8	n/a
WellChrom UV Detector K-2500 ⁸⁾	n/a	CC 2.6	n/a	CG 2.8	n/a
WellChrom UV Detector K-2501 ⁶⁾	n/a	CC 2.6	n/a	CG 2.8	n/a
WellChrom UV Detector K-2600 ^{22), 23)}	OLEE A.01.05 / A.04.05 Driver 2026	CC 2.6	Cm 6.8 Generic driver ⁵⁾	CG 2.55	n/a

Detectors (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromleon™ CDS version Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
WellChrom PDA Detector K-2700 ^{9), 23)}	n/a	PCI with RS-422 (PCI card driver supports up to Windows 7 32-bit) / LAN	n/a	CG 2.55, LAN from 3.1	n/a
WellChrom PDA Detector K-2800 ^{9), 23)}	n/a	PCI with RS-422 (PCI card driver supports up to Windows 7 32-bit) / LAN	n/a	CG 2.8, LAN from 3.1	n/a
WellChrom RI Detector 2300	n/a	CC 2.6	n/a	CG 2.8	n/a
WellChrom RI Detector 2301	n/a	CC 2.6	n/a	CG 2.8	n/a
WellChrom RI Detector 2400	n/a	CC 2.6	n/a	CG 2.8	n/a
WellChrom RI Detector 2401	n/a	CC 2.6	n/a	CG 2.8	n/a
Shimadzu fluorescence detector RF-10Axl; onboard COM or PCI/PCle COM board required ¹⁰⁾	n/a	CC 2.6	n/a	CG 3.1	n/a
Shimadzu fluorescence detector RF-20A/Axs serial control with special firmware version 0.90 only from KNAUER (<u>limited functionality</u>); onboard COM or PCI/PCle COM board required ¹¹⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.7 Driver 3055	n/a	CG 3.3.2; Driver 1787	Version 5.08.024 ²⁴⁾

Detectors (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromeleon™ CDS version Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
Shimadzu fluorescence detector RF-20A/Axs with CBM-20A only (CBM-40 does not work with the detector only; Shimadzu pump is also required)	Shimadzu Control with official Shimadzu driver; requires special license, driver not developed by KNAUER ¹⁹⁾	CC 7.2.0, only LC-10 functionality, driver developed by DataApex	Cm 7.2 with official Shimadzu driver; requires special license, driver not developed by KNAUER ²⁰⁾	n/a	n/a
Virtual Detector	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.6.5 Driver 1079	n/a	CG 3.1.6	n/a
BlueShadow Detector 10D ³⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.7 Driver 3055	n/a	CG 3.3.2 Driver 1787	Version 5.07.039 ²⁷⁾
BlueShadow Detector 40D ^{3), 7)}	OLEE A.01.05 / A.04.05 driver 1022	CC 5.0.2 Driver 4095	n/a	n/a	n/a
BlueShadow Detector 50D ^{3), 7), 23)}	OLEE A.01.05 / A.04.05 driver 1022	CC 5.0.2 Driver 4095	n/a	n/a	Version 6 R2-2 Version 5.09.134 ^{1), 27)}
Alltech Conductivity Detector 650	n/a	n/a	n/a	CG 3.1.7 Driver 446	n/a
Kontron UV Detector 332/335 ^{8), 12)}	n/a	supported by DataApex; CC 5.0.2	-	CG 3.1.7 Driver 446	n/a
Kontron UV Detector 430/430A/432/435 ^{8), 12)}	n/a	supported by DataApex; CC 5.0.2	-	CG 3.1.7 Driver 446	n/a
Kontron UV Detector 530/532/535 ^{8), 12)}	n/a	supported by DataApex; CC 5.0.2	-	CG 3.1.7 Driver 446	n/a
Kontron DAD 540(/545) ^{12), 23)}	n/a	supported by DataApex; CC 5.0.2	-	CG 3.1.7 Driver 1466	n/a

Detectors (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromeleon™ CDS version Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
Sedere Sedex ELSD 80/85LT	driver from Sedere ²⁸⁾ driver from MayLab Analytical ¹³⁾	supported by DataApex from CC 3.0.5, CC 7.4 recommended ³⁰⁾	Cm 7.2: driver from Sedere ³²⁾	supported, driver was developed by MayLab Analytical ¹⁴⁾	n/a
Sedere Sedex ELSD 90LT	driver from Sedere ²⁸⁾ driver from MayLab Analytical ¹³⁾	supported by DataApex from. CC 5.0.2, CC 7.4 recommended ³⁰⁾	Cm 7.2: driver from Sedere ³²⁾	supported, driver was developed by MayLab Analytical ¹⁴⁾	n/a
Sedere Sedex ELSD 100LT	driver from Sedere ²⁹⁾ driver from MayLab Analytical ¹³⁾	supported by DataApex from CC 5.0.2, CC 7.4 recommended ³¹⁾	Cm 7.2: driver from Sedere ³³⁾	n/a	n/a
Sedere Sedex LC	driver from Sedere ²⁹⁾ driver from MayLab Analytical ¹³⁾	supported by DataApex from CC 5.0.2, CC 7.4 recommended ³¹⁾	Cm 7.2: driver from Sedere ³³⁾	n/a	Version 5.07.039 ²⁴⁾
Sedere Sedex FP	driver from Sedere ²⁹⁾	supported by DataApex from CC 7.4 ³¹⁾	Cm 7.2: (driver from Sedere ³³⁾	n/a	n/a
ERC RefractoMax 520 series	n/a	supported by DataApex; CC 6.1.0, RS-232	Cm 7.2 (driver not developed by KNAUER)	n/a	n/a
Antec Leyden LC-EC Detector Decade	n/a	supported by DataApex; CC 5.0.2	Cm 6.8 (driver not developed by KNAUER)	n/a	Version 5.07.039

Detectors (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromeleon™ CDS version Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
Antec Decade II	n/a	supported by DataApex; CC 5.x, RS-232	-	n/a	n/a
Antec Decade Elite	n/a	supported by DataApex; CC 5.X, USB, LAN	Driver available from Antec	n/a	n/a
Shimadzu CDD-10Avp ²⁶⁾	n/a	n/a	n/a	n/a	n/a
Grace/Alltech ELSD 3300	-	supported by DataApex; CC 5.x, LAN	-	Separate, license-free driver upon request or from Grace webpage, driver not developed by KNAUER	-
Berthold HERM LB500 (also with fLumo)	n/a	Ruby script available upon request, USB	Cm 7.2: driver not developed by KNAUER, USB ³⁴⁾	n/a	n/a
Berthold Flowstar	n/a	n/a	Cm 7.2: driver not developed by KNAUER, USB ³⁴⁾	n/a	n/a
Elysia-raytest radioactivity detectors and monitors Gabi Nova, Gabi *, POMO, Ramona *	-	only upon request to Elysia; no driver from KNAUER available	-	n/a	n/a

- 1) - only LAN is supported
- 2) - to set a static IP address a separate tool is required
- 3) - via RS-232 the maximum data rate is limited to 10 Hz
- 4) - only in ClarityChrom® 5.0.2: Runs with data rates >10 Hz may not be stored with 3D data; recommended maximum data rate is 10 Hz, fixed in later versions from 5.0.3
- 5) - generic drivers do not support all device features; only control via RS-232, **data acquisition requires A/D converter**; should also work with Chromeleon™ 7.x
- 6) - a trigger on Start input also starts the internal wavelength program
- 7) - for control via RS-232 only baud rate 115200 is supported

- 8) - device only supports data acquisition via RS-232, no wavelength control
- 9) - LAN only via separate LAN adapter A4442; default communication adapter is PCI interface card for RS-422, PCI interface card requires a PCI slot on the computer; driver for PCI interface card only works with Windows NT/2000, Windows XP or Windows 7 32-bit. Older ISA interface card is not supported anymore.
- 10) - RS-232 only, onboard COM port or PCI/PCIe extension cards are recommended, USBtoCOM adapter and other adapters may not work; maximum data rate 0.5 Hz, A/D converter for higher data rates required; special RS-232 cable required (KNAUER part number G0743)
- 11) - RS-232 only, onboard COM port or PCI/PCIe extension cards are recommended, USBtoCOM adapter and other adapters may not work; special firmware version 0.90 required, feature set of RF-10 AXL only (1 channel only, no flow cell temperature control, no scan), maximum data rate 5 Hz via RS-232, A/D converter for higher data rates required; special RS-232 cable required (KNAUER part number G0743); detector sensitivity is lower compared with CBM-connection
- 12) - special RS-232 cable is required
- 13) - data rate limited to 10 Hz, driver has to be installed separately and is protected with a separate USB dongle, detector cannot be used for fractionation with KNAUER FRC option; KNAUER part number A2618-01
- 14) - data rate limited to 10 Hz, driver has to be installed separately and is protected with a separate USB dongle, detector cannot be used for fractionation with KNAUER FRC option; KNAUER part number A1513-07
- 15) - 4 channels, up to 50 Hz data rate
- 16) - 8 channels, up to 10 Hz; PurityChrom® offers only limited 3D functionality, refer to the PurityChrom® manual
- 17) - up to 2 Hz
- 18) - only one of these devices can be used in a system with PurityChrom® 5.x: RID 2.1L, CM 2.1S or Runge Mikron 81
- 19) - requires separate license, KNAUER part number A2618-02, detector cannot be used for fractionation with KNAUER FRC option
- 20) - requires separate license, KNAUER part number A1783-3
- 21) - requires separate license, KNAUER part number A1783-4
- 22) - for full functionality the KNAUER RS-232 Y-cable is required (KNAUER article number A0755)
- 23) - for acquiring 3D data (spectra) a 3D UV (PDA) license option is required.
- 24) - only one of these devices can be used in a system with PurityChrom®: IFU 2.1/IFU 2.1LAN, Sedex LC or RF-20A/Axs
- 25) - extended dynamic range for RID 2.1L is not supported in PurityChrom
- 26) - the Shimadzu CDD-10Avp requires an A/D converter for data acquisition, device control is not possible
- 27) - in PurityChrom 5.x it is not possible to run UVD 2.1S, UVD 2.1L, MWD 2.1L and 50D in any combination; only one of these device types are allowed in a system (also more than one device of the same type)
- 28) - driver from Sedere, KNAUER part number A2618-12; minimum firmware requirements: Sedex 80LT: 1.4, Sedex 85LT: 2.8, Sedex 90LT: 1.0; driver registration at Sedere is required after installation; detector cannot be used for fractionation with KNAUER FRC option
- 29) - driver from Sedere, KNAUER part number A2618-13; minimum firmware requirements: Sedex 100LT: A.1.0, Sedex LC: A.1.1, Sedex FP: A.1.1; driver registration at Sedere is required after installation; detector cannot be used for fractionation with KNAUER FRC option
- 30) - driver from DataApex, recommended minimum software version ClarityChrom 7.4, bug in version 8.5 was fixed in 8.7; driver included in installer; minimum firmware requirements: Sedex 80LT: 1.3, Sedex 85LT: 2.7, Sedex 90LT: 1.0

- 31) – driver from DataApex, recommended minimum software version ClarityChrom 7.4, bug in version 8.5 was fixed in 8.7; driver included in installer; minimum firmware requirements: Sedex 100LT: A.1.0, Sedex LC: A.1.5, Sedex FP: A.1.5
- 32) – driver from Sedere, KNAUER part number A1783-4; requires timebase Class3 license; minimum firmware requirements: Sedex 80LT: 1.4, Sedex 85LT: 2.8, Sedex 90LT: 1.0; driver registration at Sedere is required after installation
- 33) – driver from Sedere, KNAUER part number A1783-5; requires timebase Class3 license; minimum firmware requirements: Sedex 100LT: A.1.0, Sedex LC: A.1.5, Sedex FP: A.1.5; driver registration at Sedere is required after installation
- 34) – requires timebase Class3 license, KNAUER part number A1783-7

Assistants (up to three devices in one housing) ¹⁾ Included valves as fraction collector - see Fraction Collectors	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromleon™ CDS version Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
ASM 2.1L ^{1), 2)}	OLEE A.01.05 / A.04.05 Driver 1017 ³⁾	CC 3.0.7; Driver 3078 ^{4), 5), 6)}	Cm 7.2 Regular Class 3 driver 1056 ⁸⁾	CG 3.3.2; Driver 2825 ⁶⁾	Version 6 R1-2 ⁹⁾ Version 5.07.039 ^{7), 9)}
ASM 2.2L ^{1), 2)}	OLEE A.04.06 Driver 2184 ¹⁰⁾	CC 8.2.2; Driver 6036 ^{4), 6)}	Cm 7.2 Regular Class 3 driver 1056 ⁸⁾	n/a	Version 6 R1-2 ⁹⁾ Version 5.09.036 ^{7), 9)}

- 1) – configuration depends on order; refer the allowed configuration and restrictions in our FAQ (<https://www.knauer.net/en/Support/FAQ/Assistants>)
- 2) – only one LAN port for all devices
- 3) – only one valve as fraction collector, no cascading if controlled as ASM; all included modules also can be controlled as separate devices via IP port; FRC valve cascading only is possible if a valve in the ASM is configured as KNAUER MultiValve FC and is controlled as a separate device via IP port; more information about control as separate device via IP port in software manual
- 4) – ASM is only supported as complete Assistant; included modules cannot be controlled as separate devices (IP port separation is not supported)
- 5) – two pumps can be controlled either as HPG or as separate pumps from hotfix 5043, in older versions only as HPG
- 6) – only one valve as fraction collector, no cascading

- 7) - ASM is not supported as device; the included modules can be controlled as separate devices via IP port; only one valve as fraction collector
- 8) - ASM is not supported as device; the included modules can be controlled as separate devices via IP port; no valve as FRC
- 9) - ASM's Event Outs and Start In are not supported
- 10) - Modules of ASM 2.2L has to be configured as separate devices

Autosamplers (one autosampler in a system)	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromeleon™ CDS version Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
AZURA® Autosampler AS 6.1L ^{1), 2)}	OLEE A.01.05 / A.04.05 Driver 2109	CC 7.2.0.5043 ⁶⁾	Cm 7.2 Regular Class 3 driver 1038 ⁶⁾ (requires FW 1.17 or higher)	n/a	Version 6 R2-2 Version 5.08.024 ¹⁰⁾
Liquid Handler LH 2.1 (can be used as an autosampler and fraction collector)	n/a	n/a	Cm 7.2 Regular Class 3 driver, requires also Cm FRC license - currently not released	n/a	Version 5.9.149
Smartline Autosampler 3950 ^{1), 2)}	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.4.4 / LAN from CC 2.6.5 Driver 1083 ⁶⁾	Cm 7.2 Regular Class 3 driver 1024 (requires FW 1.15 or higher) Cm 6.8 SR10 Regular Class 3 driver ⁶⁾	CG 3.1.7 Driver 1466 / LAN from CG 3.3.1 Driver 1640 ⁸⁾	Version 5.07.067 ¹⁰⁾

Autosampler (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromeleon™ CDS version Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
PLATINblue Autosampler AS-1 ^{1), 2)}	OLEE A.01.05 / A.04.05 Driver 1017	CC 6.1.0, Driver 5015	Cm 6.8 SR10 Regular Class 3 driver	CG 3.3.2 PLATINblue Edition 1751 ⁸⁾ ; Regular 2825 ⁸⁾	n/a
Autosampler KNAUER Optimas	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.7 Driver 3055 ³⁾	n/a ⁹⁾	n/a ⁷⁾	Version 5.08.034
Smartline Autosampler 3800 (also WellChrom K-3800, Spark Basic Marathon)	n/a	CC 2.4.4 ³⁾	Cm 6.8; Generic driver ⁴⁾	CG 3.3.2 Driver 1787	Version 5.07.039
Smartline Autosampler 3900 (also Spark Midas)	n/a	CC2.4.4 ³⁾	Cm 6.8; Generic driver ⁴⁾	CG 2.8	Version 5.07.039
Spark Triathlon	n/a	CC 2.4.4 ³⁾	Cm 6.8; Generic driver ⁴⁾	CG 2.55	Version 5.07.039
Kontron Autosampler 360	n/a	CC 5.0.2 ³⁾	-	CG 3.1.7 Driver 1466 ⁵⁾	n/a
Kontron Autosampler 46x	n/a	-	-	CG 3.1.7 Driver 1466 ⁵⁾	n/a
Kontron Autosampler 56x	n/a	CC 5.0.2 ³⁾	-	CG 3.1.7 Driver 1466 ⁵⁾	n/a

¹⁾ - some device features have been added from time to time; therefore, not all current features are supported in the software from beginning as well as not all autosampler firmware versions support it:

AS 6.1L

- FW 1.22 - additional control options for µl-pickup (currently not supported by drivers)
- FW 1.21 - 1000 µl syringe supported (only CC: 7.2.0.5043, PC: 5.9.149), User Defined Program supported (OLEE: driver 2197, CC: 8.7 driver 6076)
- FW 1.20 - no driver-dependent changes (extra needle dry step, other needle height during first transport aspiration)
- FW 1.19 - 100 µl syringe supported: OLEE: driver 2165, CC: 8.1 driver 6034, CG: AS 6.1L not supported, Cm: n/a, PC: 5.9.149

- FW 1.18 - 30 vial tray, 2500 µl syringe can be selected also if prep mode is not enabled: OLEE: driver 2143, CC: 7.4.2 driver 5066, CG: AS 6.1L not supported; Cm: 7.x driver 1056; PC: 5.9.149
- Advanced needle wash - requires FW 1.18: OLEE: driver 2158; CC: 8.1 driver 6018; CG: AS 6.1L not supported; Cm: not supported

AS3950

- FW 1.18 - 2500 µl syringe selectable also if prep mode is not active, 30 vials tray: OLEE: driver 4.2.0.2143; CC: 7.4, driver 5066; Cm: 7.x driver 1056; CG: n/a; PC: 05.09.028 - 2500 µl syringe; 5.09.069 - 30 vial tray; 30 vial tray selectable in prep mode: OLEE: 4.2.0.2159; CC: 8.0.1.6012; Cm: 7.x driver 1056; CG: n/a; PC: 5.09.069
- FW 1.16 - 108 vial tray (Q1 2017); drivers set also 250 µl syringe and 108 vial tray as default in new method: OLEE: driver 2109; CC: 7.2 driver 5045; Cm: 7.x driver 1038; CG: n/a; PC: 05.08.034
- FW 1.15 - 0.1 µl increments, tray thermostatted up to 40 °C (2014): OLEE: A.04.05 driver 1028; CC: 5.0.3 driver 4102, Cm: 7.x, driver 1024; CG: 3.3.2 driver 2846 (40 °C only); PC: 5.07.067
- FW 1.04 - 84+3 vial tray (Aug 2011): OLEE: driver 1017; CC: 3.0.7 driver 3055; Cm: 7.x driver 1024; CG: 3.3.2 driver 1787; PC: 5.07.067
- TCP/IP support (is required to control more than one autosampler in a network, devices sold from Q3 2016 with TCP by default, former devices can be switched to TCP): OLEE: driver 1017; CC: 6.1.0 driver 5028 (hotfix); Cm: 7.x driver 1024; CG: n/a (not supported); PC: 5.07.067 (UDP not supported)
- LAN connection (devices with LAN connector sold from 2009): OLEE: driver 1017; CC: 2.6.5 driver 1094; Cm: 7.x driver 1024; CG: 3.3.1 driver 1640; PC: 5.07.067

AS-1

- FW 1.03 - 500 µl syringe is supported
- FW 1.02 - Tray thermostatted up to 40 °C: OLEE: from first version; CC: 6.1.0 driver 5015; Cm: 6.8 only; CG: 3.3.2 driver 2846; PC: AS-1 not supported

- 2) - LAN default configuration is LAN via DHCP, static IP address requires access to built-in LAN module via browser or ASM software
- 3) - driver was developed by DataApex
- 4) - upon request - driver was not developed by KNAUER
- 5) - special RS-232 cable is required
- 6) - no Stacked Injection supported
- 7) - may run with Midas/AS3900 driver
- 8) - in ChromGate TCP/IP communication is not supported, device has to be set to UDP; with UDP only one autosampler via LAN in the network can be controlled; later added features (see ¹⁾) are not supported
- 9) - generic driver for autosampler Spark Midas may work; upon request - driver was not developed by KNAUER
- 10) - Stacked Injection supported from version 5.9.149

Analog/Digital Converters	OpenLab® CDS EZChrom Edition CDS version and Driver Version ¹⁾	ClarityChrom® CDS version and Driver Version	Chromeleon™ CDS version and Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
KNAUER Interface Box IFU 2.1 LAN ²⁾	OLEE A.01.05 / A.04.05 Driver 2092	CC 6.0.1 Driver 5043 (compatible with CC/CCP 5.x versions) ¹⁵⁾	n/a	n/a	Version 6 R1-2 ⁶⁾ Version 5.08.024 ⁶⁾
KNAUER Interface Box IFU 2.1 ^{3), 4), 5)}	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.7 Driver 3078, in former version as IF2 ^{5), 15)}	n/a	CG 3.3.2 Driver 2825, in former version as IF2 ⁵⁾	Version 5.07.039 ⁶⁾
KNAUER Interface Box IF2 ⁴⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.4.4 / 2.6 ^{7), 8), 15)}	n/a	CG 3.1 ⁹⁾	Version 5.07.039 ⁶⁾
Smartline Manager 5000 ⁴⁾	n/a	CC 2.4.4/2.6 ^{7), 8), 15)}	n/a	CG 3.1 ⁹⁾	n/a
Smartline Manager 5050 ⁴⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.6.5 Driver 1100 ^{5), 15)}	n/a	CG 3.3.2 Driver 1751 ⁵⁾	n/a
KNAUER HPLC Box Model 96 ¹⁰⁾	n/a	CC 2.6 ¹¹⁾	n/a	CG 2.55 ^{12), 13)}	n/a
Kontron Interface Card ¹⁴⁾	n/a	-	-	CG 3.1.7 Driver 1466	n/a

- 1) - A/D converter in OpenLab® has to be configured separately; menu is available via *Start - Programs - Agilent Technologies - OpenLAB® CDS EZChrom Edition*, right click on *Agilent SS420X Interface Configuration* and select **Run as Administrator**
- 2) - to set a static IP address a 3rd party tool is required
- 3) - built-in USB-to-serial converter requires separate installation of a driver in Windows operating system, can be found on the HPLC software CD or KNAUER webpage (*Download* section for IFU 2.1, part number AZB00); check Windows Device Manager for the COM port number of the converter, this number has to be selected in the software configuration
- 4) - as the required serial number in the configuration, enter the last 5 digits of the serial number on the label; if the first digit is a 0 (zero), only enter the last four digits
- 5) - in previous software versions IFU 2.1 can be controlled as IF2 or Manager 5000 if IFU 2.1 is not listed in the module list
- 6) - analog input only (for data acquisition), not in a system with Sedex LC or RF-20A/Axs in PurityChrom®
- 7) - in ClarityChrom® version 2.4.4 only data acquisition via analog input is supported (max. 10 Hz), analog out supported from ClarityChrom® version 2.6
- 8) - 50 Hz option for channel 1 supported from version ClarityChrom® version 2.6

- 9) - 50 Hz option for channel 1 supported from version ChromGate® version 3.1.4
- 10) - as the required serial number in the configuration, enter the last 5 digits of the serial number on the label, if the serial number is higher than 40000, please enter only the last four digits
- 11) - supported as "Basic Edition", ClarityChrom® will switch the box into the required mode
- 12) - supported as "Integration Box" only for data acquisition via analog input; box has to be switched into the correct mode with special software
- 13) - 100 Hz option for channel 1 supported from version ChromGate® version 3.1
- 14) - requires an ISA slot on the computer; is not supported from Windows Vista (and higher)
- 15) - digital output (one per channel) cannot be programmed; is used automatically for AZ (autozero) signal

Thermostats/Column Ovens	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromleon™ CDS version Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
Column Thermostat CT 2.1	OLEE A.01.05 / A.04.05 Driver 1027	CC 5.0.3 driver 4102	Cm 7.2 Regular Class 3 driver 1024	n/a	Version 6 R2-2 ⁵⁾ Version 5.09.069 ^{4), 5)}
Support of modified CT 2.1 with S/N FCC...	OLEE A.04.06 Driver 2184	CC 8.2.3 driver 6042	Cm 7.2 Regular Class 3 driver 1056	n/a	Version 6 R2-2 ⁵⁾ Version 5.09.134 ⁵⁾
PLATINblue Column Thermostat T-1	OLEE A.01.05 / A.04.05 Driver 1017	CC 5.0.2 Driver 4095	Cm 6.8 SR10 Regular Class 3 driver	CG 3.3.2 PLATINblue Edition 1751; Regular 2825	n/a
Smartline Column Oven ²⁾ / Column Thermostat JetStream ^{1), 2)} / PLATINblue T-1 Basic ²⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.4.4	Cm 6.8 SR10 Regular Class 3 driver	CG 3.1	Version 5.07.039 ⁴⁾

Thermostats/Column Ovens (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromeleon™ CDS version Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
Smartline Column Oven 4050 ³⁾	n/a	CC 2.4.4	n/a	CG 3.1.7 Driver 1466	n/a
Memmert UN 55 Oven	n/a	n/a	n/a	n/a	Version 5.09.069 ⁴⁾
Eluent heating device ⁶⁾	n/a	n/a	n/a	n/a	Version 5.08.035 ⁴⁾

- ¹⁾ - firmware version b1.17 or higher for JetStream required
- ²⁾ - if connected, control panel has to show "Choose Function" before the instrument will be opened
- ³⁾ - device does not support static IP address, only LAN via DHCP from firmware version 1.11
- ⁴⁾ - in PurityChrom® version older than 5.09.134 only one thermostat or eluent heater is supported in one system
- ⁵⁾ - leak detector is not supported
- ⁶⁾ - separate software TempLogger for Standalone operation available

Electrical Valve Drives <u>Valves as fraction collector -> Fraction Collectors</u>	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromeleon™ CDS version Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
AZURA® Valve Unifier VU 4.1 ¹⁾	OLEE A.01.05 / A.04.05 Driver 2138	CC7.4.1 Driver 5066	Cm 7.2 Regular Class 3 driver 1056 ^{2), 3)}	n/a ¹⁾	Version 6 R1-2 Version 5.07.039
AZURA® Valve V 2.1S ¹⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 3.0.7 Driver 3078 ¹⁾	Cm 7.2 Regular Class 3 driver 1056 ^{2), 3)}	n/a ¹⁾	Version 6 R1-2 Version 5.07.039

Electrical Valve Drives (continued) <u>Valves as fraction collector -> Fraction Collectors</u>	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromleon™ CDS version and Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
WellChrom electric valve drives K-6, K-12, K-16	n/a ⁴⁾	CC 2.4.4	CM 6.8 Generic driver ³⁾	CG 2.8	Version 6 R1-2 Version 5.07.039
Smartline valve drives S6, S12, S16 ¹⁾	OLEE A.01.05 / A.04.05 Driver 1017	CC 2.4.4	CM 6.8 Generic driver ³⁾	CG 3.1 ¹⁾	Version 6 R1-2 Version 5.07.039
VICI Valco drives	OLEE A.01.05 / A.04.05 Driver 1017	CC 5.0.2 ⁵⁾	Implemented (not by KNAUER)	Generic Driver ³⁾	Version 6 R1-2 Version 5.07.039 ⁶⁾

- ¹⁾ - can be controlled as KNAUER switching valve if not listed in the module list of previous software versions (limited on RS-232 and only valve types available for V 2.1S)
- ²⁾ - generic driver for KNAUER switching valves will also work with AZURA® VU 4.1 (only for valve types available for V 2.1S, RS-232 only) and V 2.1S (RS-232 only), Cm 6.8 and 7.2
- ³⁾ - generic drivers do not support all of the device's features (e.g., RS-232 only, limited number of valve types)
- ⁴⁾ - can be controlled as KNAUER switching valve
- ⁵⁾ - driver was developed by DataApex
- ⁶⁾ - up to 8 valves supported in one system

Other Supported Instruments	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromeleon™ CDS version and Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
Advion expressIION-S CMS MS ^{1), 2)}	n/a	ClarityChrom 5.0.3	n/a	n/a	n/a
Other Supported Instruments (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver Version	ClarityChrom® CDS version and Driver Version	Chromeleon™ CDS version and Driver version	ChromGate® CDS version and Driver Version	PurityChrom® CDS and Driver version
Microsaic 4000 MiD Single Quad MS ^{1), 2)}	n/a	ClarityChrom 7.4.1	-	n/a	Version 5.09.069
Airsensor ³⁾	n/a	n/a	n/a	n/a	Version 6 R1-2 Version 5.07.039
Pressure control (pressure difference) ⁴⁾	n/a	n/a	n/a	n/a	Version 6 R1-2 Version 5.07.039
Flowmeter GJC, Model 5025000 230V	OLEE A.01.05 / A.04.05 Driver 1017	n/a	n/a	supported	n/a
Flowmeter Bronkhorst Mini Cori-Flow M12, M13, M14	OLEE A.01.05 / A.04.05 Driver 1017	CC 8.1 driver 6018	n/a	supported	Version 6 R1-2 Version 5.09.028
AZURA pH Monitor pH 2.1S	n/a	ClarityChrom 9.0	n/a	n/a	Version 6 R1-2 Version 5.07.039
Ethernet Eventbox-Interface (digital inputs and outputs)	n/a	n/a	n/a	n/a	Version 5.09.149

- 1) - driver was developed by DataApex
- 2) - device control requires MS license
- 3) - communication via game port adapter
- 4) - sensors may require KNAUER IFU 2.1/ IFU 2.1 LAN, note the different models

Fraction Collectors (one fraction collector in a system, no combination of a fraction valve and a fraction collector possible)	OpenLab® CDS EZChrom Edition CDS version and Driver version *)	ClarityChrom® CDS version and Driver version **) <small>17)</small>	Chromeleon™ CDS version Driver version ***); FRC license required	ChromGate® CDS version and Driver version ***); FRC license required	PurityChrom® CDS and Driver version
KNAUER Valve as Fraction Collector (e.g. as „MultiValve FC“, all KNAUER electric valve drives and VU 4.1)	OLEE A.01.05 / A.04.05 Driver 1017 ¹⁾	CCP 2.6 ^{1), 2)}	Cm 7.2 Regular Class 3 driver 1056; Refer Knauer FRC manual for Chromeleon	CG 3.1.6 Driver 423 ¹⁾	Version 6 R1-2 ³⁾ Version 5.07.039 ³⁾
ASM 2.2L valve	Coming soon	CC 8.2.2 6036 ¹⁴⁾	n/a	n/a	Version 6 R1-2 ¹⁶⁾ Version 5.09.036 ¹⁶⁾
ASM 2.1L valve	OLEE A.01.05 / A.04.05 Driver 1017 ¹³⁾	CCP 3.0.7 Driver 3078 ¹⁴⁾	n/a	CG 3.3.2; Driver 2825 ¹⁵⁾	Version 6 R1-2 ¹⁶⁾ Version 5.07.039 ¹⁶⁾
Liquid Handler 2.1 (can be used as an autosampler and fraction collector)	n/a	n/a	Cm 7.2 Regular Class 3 driver, requires also Cm FRC license - currently not released	n/a	Version 5.9.149
Smartline Fraction Collector 3050 ⁴⁾	n/a	CCP 2.6	n/a	CG 3.3.1 Driver 1640	n/a
Buechi Fraction Collector C-684 ⁵⁾	n/a	n/a	n/a	CG 3.1.6 Driver 423	n/a
Buechi Fraction Collector C-660	OLEE A.01.05 / A.04.05 Driver 1017	CCP 2.6	n/a	CG 3.1.6 Driver 423	n/a
ISCO Fraction Collector Foxy Jr.	n/a	CCP 2.6	Implemented (not by KNAUER)	CG 3.1.6 Driver 423	Version 5.07.039

Fraction Collectors (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver version *)	ClarityChrom® CDS version and Driver version **), ***)¹⁷⁾	Chromeleon™ CDS version Driver version; FRC license required	ChromGate® CDS version and Driver version; FRC license required	PurityChrom® CDS and Driver version
ISCO Fraction Collector Foxy R1 ⁶⁾	OLEE A.01.05 / A.04.05 Driver 1017 ⁷⁾	CCP 2.6.5 Driver 1094 (RS-232 only) ⁸⁾ CC 7.4.2 hotfix 5075 **) also LAN	Implemented (not by KNAUER)	CG3.3.2 Driver 1751 ⁷⁾	Version 6 R1-2 Version 5.07.039
ISCO Fraction Collector Foxy R2 ⁶⁾	OLEE A.01.05 / A.04.05 Driver 1017 ⁷⁾	CCP 2.6.5 Driver 1094 (RS-232 only) ⁸⁾ CC 7.4.2 hotfix 5075 **) also LAN	CM 6.8 Generic driver ²⁰⁾	CG3.3.2 Driver 1751 ⁷⁾	Version 6 R1-2 Version 5.07.039
ISCO Fraction Collector Foxy 200 ⁹⁾	n/a	CCP 2.6 and CCP 2.6.5 Driver 1100 ⁹⁾	Implemented	CG 3.1.7 Driver 1466 - CG3.3.2 Driver 1751 ⁹⁾	n/a
Virtual Fraction Collector	OLEE A.01.05 / A.04.05 Driver 1017	CCP 2.6	n/a	CG 3.1.6 Driver 423	n/a
Labocol Vario 2000 Fraction Collector	OLEE A.01.05 / A.04.05 Driver 1017	n/a	n/a	CG 3.1.6 Driver 423	Version 5.07.039
Labocol Vario 4000 Fraction Collector ⁶⁾	OLEE A.01.05 / A.04.05 Driver 1017 (RS-232 only) OLEE A.02.00 / A.04.06 Driver 2155 *) ¹⁸⁾ also LAN; only fixed rack types	CC 7.4.2 Driver 5075 (hotfix) ¹⁹⁾ only fixed rack types	-	CG 3.3.1 Driver 1640 (RS-232 only)	Version 6 R1-2 ²¹⁾ Version 5.07.039 ²¹⁾

Fraction Collectors (continued)	OpenLab® CDS EZChrom Edition CDS version and Driver version *)	ClarityChrom® CDS version and Driver version **), ***) ¹⁷⁾	Chromeleon™ CDS version Driver version; FRC license required	ChromGate® CDS version and Driver version; FRC license required	PurityChrom® CDS and Driver version
Labocol Vario 4000 Plus Fraction Collector ⁶⁾	OLEE A.02.00 / A.04.06 Driver 2155 *) ¹⁸⁾ only fixed rack types	CC 7.4.2 Driver 5075 (hotfix) ¹⁹⁾ only fixed rack types	-	n/a	Version 6 R1-2 ²¹⁾ Version 5.07.039 ²¹⁾
VICI Valco valve as VICI MultiValve Fraction Collector	OLEE A.01.05 / A.04.05 Driver 1017 ¹⁾	n/a from KNAUER, in ASM 2.1L as FC ³⁾ , ¹⁰⁾ valve as FC is supported by DataApex ¹¹⁾	n/a	CG 3.3.1 Driver 1640 ¹⁾	Version 6 R1-2 ³⁾ Version 5.07.039 ³⁾
Injection Module (combined module with pump and injection valve for injection of large injection volumes, allows for stacked injection) ¹²⁾	OLEE A.01.05 / A.04.05 Driver 1017	n/a	n/a	CG 3.3.1 Driver 1640	Version 6 R1-2 Version 5.07.039

*) - Due to changes in OpenLab EZChrom by Agilent, the KNAUER FRC AddOn needed to be adjusted and is not compatible to OpenLab EZChrom A.01.05 / A.04.05 from driver package version 2146. All later added functionality and drivers are not available for OpenLab EZChrom A.01.05 / A.04.05.

**) - Up to version 6.1, all KNAUER preparative functionality and fraction collector drivers are only available in ClarityChrom® *Prep*. From version 7.2., the ClarityChrom® *Prep* as a separate software is discontinued. The full functionality now is available in ClarityChrom® with the special Knauer FRC control. If an upgrade of ClarityChrom® *Prep* is required, a ClarityChrom with Knauer FRC control will be delivered (new User Code).

***) - The listed fraction collectors are supported in the KNAUER FRC module for ClarityChrom® (from version 7.4.2) or in former ClarityChrom® *Prep*. DataApex supports fraction collection with an own implementation without a separate license; it is available in ClarityChrom® as well as ClarityChrom® *Prep*.

1) - up to 32 valves in a cascade

2) - VU 4.1 as MultiValve FC fully supported from CC 7.4.1.5066, in previous versions as V 2.1S RS-232 only

3) - only one valve in PurityChrom® can be used as fraction collector, no cascading is supported

- 4) - switch between RS-232 and LAN with a switch on the devices rear panel
- 5) - requires special relay module for serial communication
- 6) - touch panel is not blocked in remote, do not touch it while the device is under software control
- 7) - LAN control from firmware version 1.01.03, LAN control support in firmware versions 1.01.09 and 1.01.11 faulty
- 8) - from CCP 2.6.5 1094 only RS-232 support, from 7.4.2 hotfix 5075 also LAN control
- 9) - support is discontinued due to several incompatible firmware versions; not all devices may work with the driver
- 10) - not all available valve types can be built-in into ASM 2.1L
- 11) - driver from DataApex available; does not support KNAUER fractionation features
- 12) - several pumps were added to the module when released; however, a pump is only supported in the injection module if the mentioned pump is also available as a separate pump; in ASM 2.1L pump and valve has to be controlled separately, refer software manual for detailed information
- 13) - only one valve as fraction collector, no cascading if controlled as ASM; FRC valve cascading only is possible if a valve in the ASM is configured as KNAUER MultiValve FC and is controlled as separate device via IP port; more information about control as separate device via IP port in software manual
- 14) - only one valve as fraction collector in ASM, no cascading
- 15) - only one valve as fraction collector, no cascading if controlled as ASM; FRC valve cascading only is possible if a valve in the ASM is configured as KNAUER MultiValve FC and is controlled as separate device via IP port
- 16) - only one valve as fraction collector, no cascading
- 17) - DataApex supports fraction collection with an own implementation; it is available in ClarityChrom® as well as ClarityChrom® *Prep*. KNAUER FRC option is available in ClarityChrom® *Prep* and, from version 7.4.1, also in ClarityChrom® with Knauer FRC license
- 18) - from driver version 2155, additionally model 4000 Plus, LAN control and trays with 24 and 39 vials are supported, now trays with 20, 24, 39, 80 and 125 vials are available, free tray definition is not supported; all racks must have same type (no mixed configuration supported); no USB control support; former drivers only support model 4000, RS-232 control and trays with 20, 80 and 125 vials
- 19) - only trays with 20, 24, 39, 80 and 125 vials are supported; free tray definition is not supported; all racks must have same type (no mixed rack configuration supported); no USB control support
- 20) - generic driver should also work with Cm 7, generic drivers do not support all of the device's features (e.g. RS-232 only)
- 21) - RS-232 and LAN control, free rack definition supported, mixed rack configuration supported, no USB control

Mobile Control (KNAUER)

Mobile Control requires minimum device firmware versions, which may differ between the Mobile Control versions. Refer to the list of supported devices and their required **minimum** firmware version below. A Mobile Control update may also require firmware updates of the controlled devices. Mobile Control download comes with a Firmware Wizard, which allows for firmware updates. The **Monitor Mode** is no longer supported for technical reasons. Please refer the Mobile Control page for supported features and other information:

https://www.knauer.net/en/Systems-Solutions/Software_solutions/Mobile-Control?lswitch=en

Mobile Control, Version 6.0.4 also with data acquisition support, supports the following devices:

- Pump P 2.1S/P 4.1S - Firmware V01.38
- Pump P 6.1L - Firmware V01.07
- Pump P 8.1L - Firmware V01.01
- Pump P 2.1L - Firmware V01.09
- Pump 80P APD30XX, APD60XX - Firmware V02.26
- Pump 80P APD20XX - Firmware V01.08
- Pump 40P APC30XX - Firmware V01.12
- Pump 40P APC40XX, APC60XX - Firmware V02.30
- Detector UVD 2.1S - Firmware V01.14
- Detector UVD 2.1L - Firmware V02.06
- Detector MWD 2.1L - Firmware V01.12
- Detector DAD 2.1L* - Firmware V01.12
- Detector DAD 6.1L* - Firmware V01.26
- Detector RID 2.1L - Firmware V01.24
- Detector 40D ADI01, ADI04 - Firmware V02.05
- Detector 50D ADJ01, ADJ11 - Firmware V02.18
- Valve V 2.1S - Firmware V05.01
- VU 4.1 - Firmware V05.68
- Assistant ASM 2.1L - Firmware V01.18
- Assistant ASM 2.2L - Firmware V01.14
- Column thermostat CT 2.1 - Firmware V01.06/V02.02
- Autosampler AS 3950 - Firmware V01.17
- Autosampler AS 6.1L - Firmware V01.22
- IFU 2.1 LAN - Firmware V1.05
- Fraction Collector Foxy R1/R2** - Firmware V01.01.11
- Fraction Collector Vario-4000/Plus** - Firmware V01.25

* - no 3D data acquisition, 4 channels only

** - only one rack type is supported by device; racks of different type cannot be combined in one device

X = A-Z

The **Monitor Mode** is no longer supported in OpenLab EZChrom® and ClarityChrom®.

EuroChrom for Windows 3.05 (KNAUER)

Note that EuroChrom was discontinued in 2008. It will not run on Windows systems newer than Windows XP. Fraction collectors are supported exclusively in EuroChrom Prep software.

Pumps

Smartline Pump 1000, Preparative Pump 1800, WellChrom K-1800, WellChrom K-1001, K-1000, K-501, K-120 (Smartline 100 as K-120 only via RS-232 without pressure sensor),

Feed Delivery (ECW Prep only): pump K-501, K-120

Detectors

Smartline UV Detector 2500, Smartline RI Detectors 2300 / 2400, Smartline DAD 2800 (PCI/LAN), WellChrom UV Detectors K-2000/2001, K-2500/2501, K-2600, K-200, WellChrom RI Detectors K-2301/2401, WellChrom DAD K-2700 (PCI/LAN), WellChrom DAD K-2800 (PCI/LAN), Virtual Detector

(WellChrom detectors 2xx0 allow only data acquisition, no control)

Autosamplers

AS 3800 / Spark Basic Marathon, AS 3900 / Spark Midas, Spark Triathlon

Analog/Digital Converters

KNAUER IF2 (also IFU 2.1 with built-in USBtoCOM converter), Smartline Manager 5000, KNAUER HPLC Box (Model 96)

Thermostats/Column Ovens

JetStream (FW b1.17 or later required)

Valves

KNAUER WellChrom K-6 / K-12 / K-16 [Smartline S 6 / S 12 / S 16 and AZURA® V 2.1S and VU 4.1 (only 2, 6, 12, 16 port valves) are compatible via RS-232],

Solvent Recycling: K-6 / K16

Fraction Collectors (ECW Prep 3.05 only)

Labocol Vario 2000, ISCO Foxy Jr, KNAUER WellChrom K-6 / K-12 / K-16 (also cascade up to 8 valves), Buechi 684 via short circuits, User Defined FC (via short circuits)

Other Software

OpenLab CDS (Agilent) - drivers for KNAUER AZURA® coming soon

OpenLab CDS ChemStation Edition (Agilent) - no drivers from KNAUER are available

LabVIEW (National Instruments) - support is discontinued, may not work with LabVIEW versions higher than 2011

Driver for AZURA® Pumps P 2.1S / P 4.1S, Smartline 100, BlueShadow Pumps 10P / 20P

HyStar™ (Bruker)

Compass HyStar™ 3.2 Plug-in: Drivers for PLATINblue system (Pump P-1, DAD PDA-1, UV Detector MW-1, Column Thermostat T-1, Autosampler AS-1, Manager M-1 (LPG valves only, interface module is not supported)

Compass HyStar™ 4 Plug-in: Driver for AZURA® DAD 6.1L

Xcalibur™ (Thermo)

Version 2.0.7 and 2.1: Driver for PLATINblue Pump P-1, UV Detector MW-1, DAD PDA-1, Autosampler AS-1, Column Thermostat T-1 and T-1 Basic and Smartline Pump S 1050, UV Detector S 2550, Autosampler S 3950 and Column Thermostat JetStream

TRADEMARKS

The company, product and service names used in this brochure are for identification purposes only. All trademarks and registered trademarks are the property of their respective owners.

© KNAUER

All rights reserved.

Version 1.21

Last update: November 28, 2023

KNAUER Wissenschaftliche Geräte GmbH
Hegauer Weg 38
14163 Berlin, Germany

Tel.: +49-(0)30-809727-0

Fax: +49-(0)30-8015010

E-mail: info@knauer.net

Internet: www.knauer.net

This document is subject to technical changes.
Please visit www.knauer.net for the latest version,
ordering details and overview of article numbers.